

Chemical Crystallography Laboratory	OUCB-CCL-QAP
Department of Chemistry and Biochemistry	Version 001
University of Oklahoma	January 4, 2013
Quality Assurance Plan	Page 1 of 6

Quality Assurance Plan

January 4, 2013

Chemical Crystallography Laboratory

Author: _____
Douglas R. Powell

Chemical Crystallography Laboratory	OUCB-CCL-QAP
Department of Chemistry and Biochemistry	Version 001
University of Oklahoma	January 4, 2013
Quality Assurance Plan	Page 2 of 6

Distribution

Douglas R. Powell, Laboratory Manager
 Robert Thomson, Chair, Faculty User's Committee
 Wai Tak Yip, Assistant Department Chair
 Laboratory copy maintained by Douglas R. Powell

Revision Record

Date	Version	Responsible Person	Description of Change
1/4/2013	QAP001	Douglas R. Powell	Initial Release

Table of Contents

Project Summary	3
Project Description	3
Quality Objectives and Criteria for Measurement of Data	4
Special Training and Certification	4
Documents and Records	5
Quality Control	5
Instrument Testing, Inspection, and Maintenance	5
Assessment and Reporting	6
References	6

Chemical Crystallography Laboratory	OUCB-CCL-QAP
Department of Chemistry and Biochemistry	Version 001
University of Oklahoma	January 4, 2013
Quality Assurance Plan	Page 3 of 6

Project Summary

Provide a single crystal chemical analysis service of small-molecule compounds to the students and researchers at the University of Oklahoma and more broadly to the citizens of the state of Oklahoma. A chemical analysis of a single crystal provides a geometrical description of the molecular and ionic species found in the crystal. Crystals are made up of repeating blocks of molecules. Because of the nature of the interaction between the radiation source and the atoms in these molecules, a model of the molecules from an “average” unit cell is produced.

This laboratory strives to produce the best models of the molecular geometry available from the crystals provided. This plan describes some of the activities carried out to assure the facility users that the best models are produced.

Project Description

The experiment involves shooting a beam of nearly monochromatic X-rays at a crystalline sample, measuring the relative intensities of the diffraction pattern, and analyzing the diffraction data to determine the location and types of atoms (the model). From these atom locations, the bonding geometry between the atoms is determined.

Our lab follows the guidelines for collecting and processing data, and refining chemical models that are set forth by the International Union of Crystallographers, IUCr, as described in the structure verification program checkcif (<http://checkcif.iucr.org/>). This verification program is utilized by the editors of all major journals to evaluate the quality of a crystal structure that is included in any submitted manuscript.

Small-molecule crystal structure analyses typically produce more data than is needed to uniquely determine the model, often producing 10-20 times as many unique data as there are variables in the model. These highly over-determined models nearly always lead to an unambiguous result. A statistical analysis of the data set is carried out as a part of each analysis. Standard uncertainties for the intensity data, the atomic parameters, and the derived geometry objects are always reported.

The crystal structure determinations conducted in this lab utilize a Bruker APEX ccd detector mounted on a 3-circle D8 goniometer with a sealed tube X-ray source, a mono-capillary collimator. The samples are routinely cooled to 100 K using a nitrogen-gas streaming, Oxford Cryostream700 low temperature device. This instrument was assembled and installed at the University of Oklahoma in 2002.

Chemical Crystallography Laboratory	OUCB-CCL-QAP
Department of Chemistry and Biochemistry	Version 001
University of Oklahoma	January 4, 2013
Quality Assurance Plan	Page 4 of 6

Quality Objectives and Criteria for Measurement of Data

Crystals are selected based upon their size and appearance. Single crystals are much easier to analyze than twinned crystals. Larger crystals tend to produce stronger (better) data sets. Experimental conditions such as counting time, thickness of slices of diffraction space, sample-to-detector distance are varied to produce the best data for the selected crystal. Typically several crystals are screened to find the crystal with the strongest diffraction pattern.

When possible, the counting time per image is adjusted to produce observable intensities at the highest scattering angles. Using this criteria to select the counting time, data will typically produce an average $I/\sigma(I)$ ratio of > 10 (the selected level of precision used in this lab). Unfortunately, not all crystals diffract to IUCr minimum limits of $\sin \theta_{\max} / \lambda > 0.6 \text{ \AA}^{-1}$. Data are still collected on these samples presuming that some description of the crystal structure is better than no description of the crystal structure.

The solid angle thickness of slices is routinely set at 0.3° so that each diffraction maxima appears on 2-3 consecutive slices (images) of diffraction space. When the samples scatter poorly producing large spots, the thickness of each image is adjusted to make 2-3 slices through each spot.

In most cases, leaving the sample-to-detector distance at 5 cm is sufficient. However, when a unit cell axial length in the sample $> 25 \text{ \AA}$, the diffraction maxima appear to blend together forming diffraction lines. The sample-to-detector distance is increased to produce diffraction patterns with discrete spots. Note that when this distance is $> 7 \text{ cm}$, then data must be collected with at least two swing angle settings to collect sufficient data for the IUCr minimum standards.

To attain the best quality measurements of our diffraction intensities, multiple measurements of the symmetry-unique data are made. For systems with monoclinic or higher symmetry, our lab tries to achieve an average data redundancy of at least 4. For triclinic systems, our lab tries to achieve an average data redundancy of at least 3.

Special Training and Certification

All users of the facility must have annual training in radiation safety from the University's Radiation Safety Office (<http://www.ouhsc.edu/rso/>), as well as annual training in chemical hazards, chemical waste disposal, and fire safety. The lab manager should have several years of experience conducting crystal structure studies, as well as training from the manufacturer on the operation and maintenance of the instrument. Other users of the facility will be trained by the lab manager in the use of the instrument and computer software.

Chemical Crystallography Laboratory	OUCB-CCL-QAP
Department of Chemistry and Biochemistry	Version 001
University of Oklahoma	January 4, 2013
Quality Assurance Plan	Page 5 of 6

Documents and Records

Records of the radiation safety training will be maintained by the University's Radiation Safety Office. Records of the chemical and fire safety training will be maintained by the University's Environmental Health and Safety Office. Records of lab users training will be maintained by the CCL laboratory manager.

Quality Control

Users are to follow the procedures outlined in the series of standard operating procedures prepared for the Chemical Crystallography Laboratory.

APEX Instrument Operating Instructions	OUCB-CCL-SOP-1.001	1/4/2013
Alignment and Calibration of the APEX Instrument	OUCB-CCL-SOP-2.001	1/4/2013
Mount Sample	OUCB-CCL-SOP-3.001	1/4/2013
Center the Sample on the Instrument	OUCB-CCL-SOP-4.001	1/4/2013
Data Collection Using SMART	OUCB-CCL-SOP-5.001	1/4/2013
Data Reduction and Processing	OUCB-CCL-SOP-6.001	1/4/2013
Structure Refinement	OUCB-CCL-SOP-7.001	1/4/2013
Structure Reporting	OUCB-CCL-SOP-8.001	1/4/2013

All crystal structures determined in this lab will be evaluated using the IUCr checkcif (<http://checkcif.iucr.org/>). Serious issues such as A and B level alerts must be addressed in a lab report describing the structure.

Final reports are prepared using a word processing program that automatically checks for spelling and grammar errors.

Instrument Testing, Inspection, and Maintenance

Weekly checks of the general laboratory conditions will be made to assure that all pieces of equipment are working as specified. Problem issues will be noted and addressed as soon as possible.

The precision of the instrument will be assessed by the results of a semi-annual analysis of an ylid sample, provided by the manufacturer. For these analyses, the instrument first will be aligned according to the procedures in Alignment and Calibration of the APEX instrument. Then intensity data will be collected and processed by a standard procedure, and the values of

Chemical Crystallography Laboratory	OUCB-CCL-QAP
Department of Chemistry and Biochemistry	Version 001
University of Oklahoma	January 4, 2013
Quality Assurance Plan	Page 6 of 6

average I, average $I/\sigma(I)$, cell parameters, and refined R_1 and wR_2 kept by the lab manager as a permanent lab record for this instrument.

Annually these preventive maintenance steps will be performed: check and clean all screens in the water cooling system for the generator, replace the cooling water in the Haskris and Neslab water coolers, clean and lubricate the gears of the omega and 2theta drives, empty drain and dry the 60 liter supply Dewar for the LT device, check all voltages in the D8 controller using the D8Tools program.

Assessment and Reporting

The checkcif report for each crystal structure will be included with data files for each project. External assessment of crystal structures submitted for publication is provided by reviewers of manuscripts that include the crystal structures.

Users are encouraged to discuss ideas for improving lab activities with either the lab manager or the faculty supervisor chairing the lab faculty committee. Annual surveys will be sent out to all users asking for additional input. All suggestions will be strongly considered and, when possible, will be implemented.

An annual report of the productivity of the facility is presented to both the chair of faculty oversight committee and the Department's assistant chairman.

References

<http://checkcif.iucr.org/>

Web site to submit crystal structure model and data to verify and validate one or more crystal structures.

<http://www.ouhsc.edu/rso/>

Web site for the Radiation Safety Office of the University of Oklahoma.

Alignment and Calibration of the APEX Instrument, SOP OUCB-CCL-2.001, Chemical Crystallography Laboratory, Department of Chemistry and Biochemistry, University of Oklahoma.